

**Abstract number:** Euro18A-POS134

**Abstract type:** Poster

**Reference:** This abstract is a part of the EuroPCR 2018 programme, 22-25 May 2018, Paris

**Link:** <https://abstractbook.pconline.com/export/pdf/id/100310>

**Published on:** 15 May 2018

## **Safety and clinical performance of biodegradable polymer-coated ultrathin everolimus-eluting stents in "real-world" patients: a multicentre registry (PERFORM-EVER)**

KASTURI S. (1), POLASA S. (2), REDDY V. K. (3), ALI M. (4), KUMAR C. (4), REDDY V. K. (4), KUMAR C. (4), SINGH S. (1), REDDY V. K. (1), SHEWALE S. (1)

(1) Sunshine Hospitals, Secunderabad INDIA(2) Chalmeda Anand Rao Institute of Medical Sciences, Karimnagar INDIA(3) OMNI Hospitals, Kurnool INDIA(4) Gowri Gopal Hospital, Kurnool INDIA

**THEME:** Coronary Interventions

**TOPIC(S):** Stents and scaffolds

### **AIMS**

To report 1-year clinical safety and efficacy outcomes for ultrathin-strut (60  $\mu$ m) biodegradable polymer coated Tetrilimus everolimus-eluting stent (EES) [Sahajanand Medical Technologies Pvt. Ltd., Surat, India] in patients with coronary artery disease in real-world clinical practice.

### **METHODS AND RESULTS**

The PERFORM-EVER was an observational, multicenter, single-arm and investigator-initiated retrospective registry. All "real-world" patients who had received Tetrilimus EES between July-2015 and October-2016 at our centers were analyzed. The data were collected retrospectively either by extraction from existing databases in consecutive fashion where index and follow-up data existed or the follow-up was obtained by telephonic contact. Primary endpoint was 1-year incidence of target lesion failure, which was defined as a composite endpoint of cardiac death, myocardial infarction, and target lesion revascularization by percutaneous or surgical methods. The Academic Research Consortium-defined stent thrombosis was assessed as additional safety endpoint. During the study period, 815 Tetrilimus EES (1.4 $\pm$ 0.5 stent/patient) were implanted to treat 735 coronary lesions (1.1 $\pm$ 0.3 stent/lesion) in 594 patients (mean age: 55.6 $\pm$ 12.1 years). Among them, 453 (76.3%) were males, 192 (32.3%) were hypertensives, 138 (23.2%) were diabetics, 205 (34.5%) were alcoholics, 141 (23.7%) were smokers, 106 (17.8%) were tobacco chewers, 46 (7.7%) had previous revascularization, and 225 (37.8%) displayed multi-vessel coronary disease. Of treated lesions, 577 (78.5%) were complex (i.e. Type B2/C) and 103 (14.0%) had total occlusion. The average length and diameter of implanted Tetrilimus EES were 27.6 $\pm$ 9.7 mm and 3.0 $\pm$ 0.3 mm respectively. Post-implantation, events of target lesion failure were reported in 5 (0.8%) patients at 30-day follow-up and 13 (2.2%) patients at 6-month follow-up. The cumulative major adverse cardiac event rate at 1-year follow-up was 3.7%, which included 9 (1.5%) cardiac deaths, 8 (1.4%) myocardial infarctions, and 5 (0.8%) target lesion revascularizations. There were 5 (0.8%) cases of probable stent thrombosis and 4 (0.7%) cases of possible stent thrombosis at 1-year follow-up.

### **CONCLUSIONS**

Low incidences of target lesion failure and stent thrombosis at 1-year follow-up indicates that Tetrilimus EES may have encouraging safety and efficacy in unselected real-world patients with coronary artery disease, including those with high-risk characteristics and complex lesions.